

**SPONSOR** 

13065

William Romey PTM&W Industries, Inc. 10640 S. Painter Ave Santa Fe Springs, CA 90670 **REPORT DATE:** 4/28/14



MSP NUMBER: 14C0332H

P.O. NUMBER: 25208

**PAGE**: 1 of 1

## CERTIFICATE

## CLASS VI - 70°C PLASTIC TESTS (USP 36 <88>, REV. 5/2013)

Name: PTM&W Industries PT8902A/B Urethane Casting System

Physical Description: Solid

Total Quantity Received for Testing: 8 pieces 60 cm<sup>2</sup>, 16 strips, 2 strips, 1 piece 1.5x1.5 inches

Part Number: PT8902A/B

Storage Condition: Room Temperature

Date Received: 3/21/14

Class VI - Intracutaneous Test

PBL Report No.: 14C0332H-X02G

Class VI - Systemic Injection Test

PBL Report No.: 14C0332H-X03G

Class VI –Intramuscular Implantation Test

PBL Report No.: 14C0332H-X04G

The test article met the requirements for Plastic Class VI - 70°C testing per USP <88>.

Multiple Study Project Coordinator

Pacific BioLabs

**QAU** Review